Faculty of Science and Engineering

School of Engineering, Computing and Mathematics

2019-06-20

Detection of Alzheimer's Protein Biomarker Clusterin using novel Graphene Biosensors

Awan, Shakil

http://hdl.handle.net/10026.1/14999

All content in PEARL is protected by copyright law. Author manuscripts are made available in accordance with publisher policies. Please cite only the published version using the details provided on the item record or document. In the absence of an open licence (e.g. Creative Commons), permissions for further reuse of content should be sought from the publisher or author.

Detection of Alzheimer's Protein Biomarker Clusterin using novel Graphene Biosensors

Shakil A. Awan

¹Wolfson Nanomaterials & Devices Laboratory, School of Computing, Electronics and Mathematics, University of Plymouth, Plymouth, Devon, PL4 8AA, UK

Email: shakil.awan@plymouth.ac.uk

Worldwide prevalence of dementia is estimated to rise from the current 50 million people affected to 152 million by 2050, costing healthcare systems ~2% of the global GDP to diagnose, treat and care for patients [1]. Although there are over 130 different types of dementias, Alzheimer's disease (AD) accounts for 60-70% of all cases. Methods for early diagnosis of AD could significantly impact disease detection, progression monitoring and therapeutics. For these applications we are currently developing novel graphene sensors that can detect protein biomarker concentration changes in serum/plasma samples (compared to samples from healthy patients) with high sensitivity and specificity [2-4]. The aim of this talk is to present a brief overview of graphene electronic properties and why they may be important for applications in protein biomarker detection. The talk will then detail our approach for ultra-sensitive detection of proteins to ~10fg/mL level, fabrication of the graphene sensors, latest results on Clusterin (and hCG cancer biomarker [5]) and future directions of the research in terms of molecular dynamics simulations and multiplexing of the graphene sensors for point-of-care applications in the near future.

We acknowledge funding from the UK's EPSRC EP/M006301/1 and University of Plymouth, GD105227.

References

- [1] World Alzheimer's Report 2018, Alzheimer's Disease International, Sept 2018.
- [2] Haslam C., Awan S.A., et al., "Label-free sensors based on graphene field-effect transistors for the detection of Human Chorionic Gonadotropin cancer risk biomarker", *Diagnostics*, **8**, 5, 1-13, 2018.
- [3] Awan S. A., Lombardo A., et al., "Transport conductivity of graphene at RF and microwave frequencies", 2D Mater., 3 (1), 015010-1-11, 2016.
- [4] Awan S. A., Kibble B. P. and Schürr J., "Coaxial Electrical Circuits for Interference-Free Measurements", Book, *Institution of Engineering and Technology*, March 2011, ISBN: 1849190690.
- [5] Damiati S., Awan S. A. et al., "Sensitivity Comparison of Macro- and Micro-electrochemical Biosensors for Human Chorionic Gonadotropin (hCG) Biomarker Detection", *submitted*. arXiv announcement 7/6/19.